

#### Towards Uniform UV-spectra (UV2NIST)


Annemieke Kolkman, Erik Emke, Albert Brandt Norman MassBank Workshop, Amsterdam, 27<sup>th</sup> November 2012



Watercycle Research Institute

# Outline

- HPLC-UV fingerprint
  - Harmonized method
  - KRretl
- UV spectrum library (UV2NIST)
  - Requirements
  - Development






Significance of HPLC/UV fingerprinting in monitoring water quality

- Monitoring organic compounds in time and place (daily at monitoring station in Bimmen and Eijsden, Keizersveer)
- Comparison of raw water, water types and products
- Assessment of water treatment
- Early-warning system
  - Unknown and known compounds
- Direction of identification of unknown substances by other techniques (TOF/Orbitrap)






# Harmonizing HPLC-UV Fingerprint method (basic requirements)

-Use a **<u>C18 column</u>** 5µm of 250 \* 4,6 mm or one with equivalent separation power

- The gradient between the internal standards is linear
- The pH is kept constant with 0,05% formic acid in both eluents
- The DAD is set at a:
  - acquisition rate of at least 1 Hz
  - with a set resolution of < 1,5 nm
  - Spectral range 200-350 nm or higher
- The "trigger" wavelength is at 215 nm



- Normalize the concentration to response of chloorxuron at 215 nm

#### *KRretI*: Retention index

$$KRretI = Rty_a = Rt_{Fn} + \frac{(Rt_{Nb} - Rt_{Fn})}{(Rt'_{Nb} - Rt'_{Fn})}(Rtx_a - Rt'_{Fn})$$


 $\begin{array}{l} \operatorname{Rt}_{Fn} \text{ and } \operatorname{Rt}_{Nb} = \text{Determined standard retention time of} \\ \operatorname{fenuron} \left( 21.12 \text{ min} \right) \text{ and neburon} \left( 43.6 \text{ min} \right) \\ \operatorname{Rt'}_{Nb} \text{ and } \operatorname{Rt'}_{Fn} = \text{Measured retention time of the two internal standards} \\ \operatorname{Rt}_{ya} = \text{Corrected retention time for the compound with retention time } \operatorname{Rt}_{Xa} \\ \operatorname{Rt}_{Xa} = \text{Actual measured retention time compound } X \end{array}$ 

## Results round robin tests 2011

	KRetl (min.)	Intralab (min.)	Interlab (min.)	
		precision (n=4)	precision (n=6)	
Caffeïne	14,98	0.29	0.02	
Fenuron	21,12	I.S	I.S	
Carbamazepine	28,05	0,10	0.01	
Methabenzthiazuron	30,76	0,04	0.01 0.01 0.01	
ТРРО	33,13	0,70		
Chloorxuron	38,24	0,11		
Linuron	38,98	0,51	0.01	
Metolachlor	42,83	0,03	0.01	
Neburon	43,60	I.S	I.S	


# Store the UV-spectra in a database

- Libraries are not easily transferred from one DAD to the other (history is lost)
- Exchangeability between DAD from different manufacturers is difficult
- Building a new library is time-consuming
- sharing data of known compounds with other participants
- New spectra from "unknown" compounds can be shared in a uniform way





## UV2NIST convertor (Step by step)

- 1. Read the output of DAD detector
  - Currently support for Agilent and Waters equipment
- 2. Correct the wavelength for instrument depended wavelength accuracy
- 3. Scale the wavelength to 1 nm
- 4. Normalize the absorbance
- 5. Calculate the *KRetl* retention index
- 6. Add a comment with the origins of the spectra
- 7. Save it in MSP format (compatible with NIST-MS search)



# Import in NIST-MS search





	UV2Ma	ssBank?					
	ex Type : ontributor: UV2NIST Sp	bectra from KWR				<u>Edit</u>	/ Resubmit Quer
Res	ults: 3 Hit. (1 - 3 Displa	ayed )			Open All Tree	Multiple Display	Spectrum Search
Firs	st Prev <b>1</b> Next Las	st (Total <b>1</b> Page)					▼ Results End
		Name	Δ		la / Structure	ExactMass	ID
	■ Atrazine		1 spectrum	C8H14CIN5	HC PH	215.09380	
	<u>UV; N/A</u>						UV000079
	■ Caffeine		1 spectrum	C8H10N4O2	H.C.	194.08040	
	<u>UV; N/A</u>						UV000054
	■ MDEA		1 spectrum	C12H17NO2	HE TO THE	207.12590	
	<u>UV: N/A</u>						UV000001


#### UV2MassBank

#### MassBank Record: UV000054

Home | Spectrum | Quick | Peak | Substructure | Browser | Browse | Index | MassBank ID:

Go

#### Caffeine; UV; N/A;



ACCESSION: UV000054 RECORD\_TITLE: Caffeine; UV; N/A; DATE: 2009.08.31 (Created 2012.11.16) AUTHORS: E. Emke: KWR, NL (modified by E. Schymanski: Eawag, CH) LICENSE: CC BY-SA COPYRIGHT: KWR COMMENT: 14.82;LB1;AP1;ONB;AW1;2009/8/31 3:32:08 COMMENT: UV2NIST2MassBank

Searching UV2MassBank								
Quick Search								
Home   Spectrum   Quick   Peak   Substructure   Browser   Browse   Index   MassBank ID:	Go							
Search by Keyword     Search by Peak								
Compound Name atrazine AND • Exact Mass Tolerance 0.3 AND • Formula (e.g. C6H7N5, C5H*N5, C5*) Reset	Instrument Type							
	MS Type All MS MS2 MS3 MS4 N/A Ion Mode Positive Negative  Both 16							

# **Final Conclusions**

- A uniform database of UV spectra has been developed
  - Instrument independent (Wasters, Agilent, Thermo,...)
  - Data sharing is possible
  - Database contains 103 compounds (DOA's, pesticides, pharmaceutical, industrial compounds)
- The database is stored in text format
- Data can be searched with free software (e.g. NIST-MSsearch demo version)
- Ready to export to other database tools (e.g. Massbank)

# Acknowlegments



